Nithsdale Wheelchairs

Applies to:

01659 490 667

Mercury Scooters (after 09/03) (Code: M36 / M46 / M48) Pioneer Scooter (Code: MS010) Vitesse 8 Scooter (Code: and MS032) Neo 8 Scooter (Code: MS042)

Diagnostics

Below is a break down of the flash sequence giving an indication of the area which is at fault. If there is a fault present the controller will indicate the fault by a series of flashes. The flashes will be spaced between short pauses then there will be a longer pause before the series starts again. Please be aware that 1 flash (low battery voltage) will be reported by the end user as constant flashing. The scooter simply requires recharging or replacement batteries.

On the Vitesse 8 and Neo 8, the status LED will flash to indicate the fault. On the Mercury range, the left-most segment of the battery gauge will flash to indicate the fault. The status LED / battery gauge will continue flashing until the fault is rectified and the keyswitch has been cycled



1 Flash •••	Low Battery. Battery voltage has dropped below 23.3V. Recharge batteries as soon as possible.
2 Flash	Battery voltage too low. Voltage has dropped below 16.5V. Recharge batteries or leave scooter switched off a few minutes.
3 Flash	Battery voltage too high. Unplug the charger from the scooter if it is plugged in. Turn the scooter off and on. Reduce speed down any gradients.
4 Flash	Current limit time-out. The scooter has drawn too much current for too long. Turn the scooter off and leave for a minute. If the problem persists, check loom and motor for shorts.
5 Flash	Brake Fault. Check the freewheel lever is in the DRIVE position. Check the brake connector is plugged in firmly.
6 Flash	Out of Neutral at Power Up. Return wigwag paddle to neutral then turn off and on again. If problems persist, wigwag paddle may need re-calibrating
7 Flash	Speed Pot Error . Wigwag paddle or its wiring is faulty. Check the paddle is calibrated and check for open or short circuits.
8 Flash	Motor Volts Error. Motor or its wiring is faulty. Check the paddle is calibrated and check for open or short circuits.
9 Flash	Internal Controller Fault. Contact Drive Medical